

Declaration of Conformity

We, Schaffner EMV AG, declare, under the sole responsibility, that the following products:

FN2010B	FN2080B	FN9233B	FN9263B	FN9290B
FN2020B	FN2090B	FN9244B	FN9264B	FN9292B
FN2030B	FN2091B	FN9246B	FN9266B	FN9299B
FN2060B	FN2092B	FN9255B	FN9280B	
FN2070B	FN9222B	FN9260B	FN9282B	
FN2071B	FN9226B	FN9262B	FN9289B	

are suitable for use in medical equipment requiring one Means of Protection (MOP) with regards to leakage currents, insulation, air clearances and creepage distances according to the following standards:

Standard	Edition	Title
IEC 60601-1	3.1: 2012-08	Medical electrical equipment - Part 1: General requirements for basic safety and essential performance

Note:

All **B**-type product versions, options, terminal and housing styles are covered under the respective product family name listed above.

Luterbach, 15.04.2020



Beat Stauffer
Head of Global R&D EMC